

Product datasheet

info@arigobio.com

ARG82800 Human Free Triiodothyronine (fT3) ELISA Kit

Package: 96 wells Store at: 4°C

Summary

Product Description ARG82800 Human Free Triiodothyronine (fT3) ELISA Kit is an enzyme immunoassay kit for the

quantitative determination of Free Triiodothyronine (fT3) concentration in Human serum.

Tested Reactivity Hu
Tested Application ELISA

Specificity Cross-Reactivity:

L-Triiodothyronine: 100% D-Triiodothyronine: 34% Triiodothyropropionic acid: 20% Diiodo-D-thyronine: 0.5%

D-Thyroxine: 0.3% L-Thyroxine: 0.9%

Target Name Free Triiodothyronine (fT3)

Conjugation HRP

Conjugation Note Substrate: TMB and read at 450 nm.

Sensitivity 0.3 pg/ml
Sample Type Serum

Standard Range 2 - 40 pg/ml

Sample Volume $25 \mu l$

Precision Intra-Assay CV: 6.7%

Inter-Assay CV: 8.2%

Application Instructions

Assay Time 1 hour 15 min

Properties

Form 96 well

Storage instruction Store the kit at 2-8°C. Keep microplate wells sealed in a dry bag with desiccants. Do not expose test

reagents to heat, sun or strong light during storage and usage. Please refer to the product user manual

for detail temperatures of the components.

Note For laboratory research only, not for drug, diagnostic or other use.

Bioinformation

Background L-Triiodothyronine, a thyroid hormone, circulates in blood almost completely bound (>99.5%) to carrier

proteins.

The main transport protein is thyroxine-binding globulin (TBG). However, only the free (unbound)

www.arigobio.com arigo.nuts about antibodies 1/2


portion of triiodothyronine is believed to be responsible for the biological action. Furthermore, the concentrations of the carrier proteins are altered in many clinical conditions, such as pregnancy. In individuals with normal thyroid function, as the concentrations of the carrier proteins change, the total T3 levels change in consert so that the free thriiodothyronine (fT3) concentration remains constant. Thus, measurements of fT3 concentrations correlate more reliably with clinical status than total triiodothyronine levels.

For example, the increase in total triiodothyronine levels associated with pregnancy, oral contraceptives, and estrogen therapy result in higher total T3 levels while the fT3 concentration remains basically unchanged. This microplate enzyme immunoassay methodology provides the technician with optimum sensitivity while requiring few technical manipulations in a direct determination of fT3

Research Area

Signaling Transduction kit

Images



ARG82800 Human Free Triiodothyronine (fT3) ELISA Kit standard curve image

ARG82800 Human Free Triiodothyronine (fT3) ELISA Kit results of a typical standard run with optical density reading at 450 nm.