



Human Herpes simplex virus 2 IgG antibody ELISA Kit

Human Herpes simplex virus 2 IgG antibody ELISA Kit has been designed for the qualitative determination of specific IgG antibodies against Herpes 2 in serum and plasma (Citrate, heparin).

Catalog number: ARG82879

Package: 96 wells

For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	Page
INTRODUCTION.....	3
PRINCIPLE OF THE ASSAY	4
MATERIALS PROVIDED & STORAGE INFORMATION	5
MATERIALS REQUIRED BUT NOT PROVIDED	5
TECHNICAL HINTS AND PRECAUTIONS	6
SAMPLE COLLECTION & STORAGE INFORMATION.....	6
REAGENT PREPARATION	7
ASSAY PROCEDURE	8
CALCULATION OF RESULTS.....	9
INTERPRETATION OF RESULTS.....	10
QUALITY ASSURANCE.....	11

MANUFACTURED BY:

Arigo Biolaboratories Corporation

Address: No. 22, Ln. 227, Gongyuan Rd., Hsinchu City 300, Taiwan

Phone: +886 (3) 562 1738

Fax: +886 (3) 561 3008

Email: info@arigobio.com

INTRODUCTION

Herpes simplex virus 1 and 2 (HSV-1 and HSV-2), also known by their taxonomical names Human alphaherpesvirus 1 and Human alphaherpesvirus 2, are two members of the human Herpesviridae family, a set of new viruses that produce viral infections in the majority of humans. Both HSV-1 (which produces most cold sores) and HSV-2 (which produces most genital herpes) are common and contagious. They can be spread when an infected person begins shedding the virus.

About 67% of the world population under the age of 50 has HSV-1. In the United States, about 47.8% and 11.9% are believed to have HSV-1 and HSV-2, respectively. Because it can be transmitted through any intimate contact, it is one of the most common sexually transmitted infections.

Many of those who are infected never develop symptoms. Symptoms, when they occur, may include watery blisters in the skin or mucous membranes of the mouth, lips, nose, or genitals, or eyes (Ocular herpes). Lesions heal with a scab characteristic of herpetic disease. Sometimes, the viruses cause mild or atypical symptoms during outbreaks. However, they can also cause more troublesome forms of herpes simplex. As neurotropic and neuroinvasive viruses, HSV-1 and-2 persist in the body by hiding from the immune system in the cell bodies of neurons. After the initial or primary infection, some infected people experience sporadic episodes of viral reactivation or outbreaks. In an outbreak, the virus in a nerve cell becomes active and is transported via the neuron's axon to the skin, where virus replication and shedding occur and cause new sores. [Provided by Wikipedia: Herpes simplex virus]

PRINCIPLE OF THE ASSAY

This assay employs the enzyme immunoassay technique. Specific antigen has been pre-coated onto a microtiter plate. Each sample or Calibrator A to C are pipetted into the wells and any specific Antibody present is bound by the immobilized antigen. After washing away any unbound substances, a HRP-conjugated anti human IgG antibody is added to each well and incubate. After washing away any unbound antibody-enzyme reagent. The immune complex formed by the bound conjugate is visualized by adding TMB substrate which gives a blue reaction product. The intensity of this product is proportional to the amount of specific antibodies in the sample. The color development is stopped by the addition of Stop Solution and the intensity of the color is measured at a wavelength of 450 nm.

Human Herpes simplex virus 2 IgG antibody ELISA Kit ARG82879

MATERIALS PROVIDED & STORAGE INFORMATION

Store the unopened kit at 2-8 °C. Use the kit before expiration date.

Component	Quantity	Storage information
Antigen-coated microplate (HSV-2 recombinant antigens)	8 X 12 strips	4°C.
Calibrator A (Negative Control)	2 ml (Ready-to-use)	4°C.
Calibrator B (Cut-off Standard)	3 ml (Ready-to-use)	
Calibrator C (Positive Control)	2 ml (Ready-to-use)	
HRP-conjugated antibody	20 ml (Ready-to-use)	4°C
20X Wash buffer	50 ml	4°C
Sample Diluent	100 ml (Ready-to-use)	4°C
TMB Substrate	15 ml (Ready-to-use)	4°C (Protect from light)
STOP Solution	15 ml (Ready-to-use)	4°C
Plate sealer	1 piece	4°C

MATERIALS REQUIRED BUT NOT PROVIDED

- Microplate reader capable of measuring absorbance at 450 / 620 nm
- Incubator 37°C
- Vortex / mixer
- Pipettes and pipette tips
- Deionized or distilled water
- Automated microplate washer (optional)

TECHNICAL HINTS AND PRECAUTIONS

- Wear protective gloves, clothing, eye, and face protection especially while handling blood or body fluid samples.
- Store the kit at 4°C at all times and do not use after the expiry date.
- It is very important to bring all reagents and samples to room temperature (20-25°C) and mix them before starting the test run.
- Unused strips should be resealed in the aluminium foil along with the desiccant supplied and stored at 2-8 °C.
- If crystals are observed in the 20X Wash buffer, warm up to 37°C until the crystals are completely dissolved. The TMB Color developing agent should be colorless (or could have a slight blue tinge) and transparent before using.
- Ensure complete reconstitution and dilution of reagents prior to use.
- All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
- Before using the kit, spin tubes and bring down all components to the bottom of tubes.
- All materials should be equilibrated to room temperature (RT; 20-25°C) before use.
- It is highly recommended that the standards, samples and controls be assayed in duplicates.
- Perform all assay steps in the order given and without any delays.
- Change pipette tips between the addition of different reagent or samples.
- For further internal quality control each laboratory should additionally use known samples. For further internal quality control each laboratory should additionally use known samples.

SAMPLE COLLECTION & STORAGE INFORMATION

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

Serum: Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately. The samples can be stored at 2-8 °C up to 5 days or aliquot and store samples at ≤ -20 °C or lower for longer storage. Avoid repeated freeze-thaw cycles.

Plasma: Collect plasma using citrate or heparin as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately. The samples can be stored at 2-8 °C up to 5 days or aliquot and store samples at ≤ -20 °C or lower for longer storage. Avoid repeated freeze-thaw cycles.

Note:

- Heat inactivation of samples is not recommended.
- Before assaying, all samples should be diluted 1+100 with Sample Diluent Buffer. Dispense 10 µL of sample and 1 mL of Sample Diluent Buffer into tubes to obtain a 1+100 dilution and thoroughly mix with a Vortex.

REAGENT PREPARATION

- **1X Wash buffer:** Dilute 20X Wash buffer into distilled water to yield 1X Wash buffer. (E.g., add 50 mL of 20X Wash Buffer into 950 mL of distilled water to a final volume of 1000 mL)

ASSAY PROCEDURE

All materials should be equilibrated to room temperature (20-25°C) before use. Standards and samples should be assayed in duplicates.

1. Remove excess microplate strips from the plate frame, return them to the foil pouch containing the desiccant pack, and reseal it.
2. Add **100 µL** of **diluted samples** and **controls** into respective wells. Leave one well empty for the **substrate blank**.
3. Cover the plate with the foil and incubate for **60 ± 5 minutes** at **37 ± 1 °C**.
4. Aspirate each well and wash, repeating the process 2 times for a total **3 washes**. Wash by filling each well with **1× Wash Buffer (300 µL)** using a squirt bottle, manifold dispenser, or autowasher. The interval between washing and aspiration should be **> 5 sec**. Complete removal of liquid at each time is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating, decanting or blotting against clean paper towels.

Note: If performing the test on ELISA automatic systems we recommend increasing the washing steps from **three** up to **five** and the volume of Washing Buffer from **300 µL** to **350 µL** to avoid washing effects.

5. Add **100 µL** of **HRP-conjugated antibody** into each well (**except the substrate blank well**). Incubate for **30 minutes** at **RT** in the dark.
6. Wash as according to step 4.
7. Add **100 µL** of **TMB Substrate** to each well (including the well for substrate blank). Cover the plate and incubate for exactly **15 minutes** at **RT** in the dark.

Human Herpes simplex virus 2 IgG antibody ELISA Kit ARG82879

8. Add **100 µL** of **Stop Solution** to each well (including substrate blank wells).
9. Read the OD with a microplate reader at **450 nm** within **30 minutes**. (**620 nm** as optional reference wave length) and use the substrate controls as blank.

CALCULATION OF RESULTS

1. Adjust the ELISA Microtiterplate reader to zero using the Substrate Blank.
If - due to technical reasons - the ELISA Microplate reader cannot be adjusted to zero using the Substrate Blank, subtract its absorbance value from all other absorbance values measured in order to obtain reliable results.
2. Measure the absorbance of all wells at 450 nm and record the absorbance values for each standard / control and sample in the plate layout. Measurement using a reference wavelength of 620 nm is recommended. Where applicable calculate the mean absorbance values of all duplicates.
3. In order for an assay run to be considered valid, these Instructions for Use have to be strictly followed and the following criteria must be met:
Substrate Blank: absorbance value < 0.100
Calibrator A (Negative control): absorbance value < 0.200 and < Cut-off
Calibrator B (Cut-off standard): absorbance value > 0.150 – 1.300
Calibrator C (Positive control): absorbance value > Cut-off
If these criteria are not met, the test is not valid and must be repeated.

INTERPRETATION OF RESULTS

1. The Cut-off is the mean absorbance value of the Cut-off standard determinations.
2. $(\text{Sample (mean) absorbance value} \times 10) / \text{Cut-off} = U$ (Units)
3. **< 9U (Negative):** The sample contains no antibodies against the pathogen. A previous contact with the antigen (pathogen resp. vaccine) is unlikely.
4. **9-11U (Equivocal):** Antibodies against the pathogen could not be detected clearly.
It is recommended to repeat the test with a fresh sample in 2 to 4 weeks.
If the result is equivocal again the sample is judged as negative.
5. **>11U (Positive):** Antibodies against the pathogen are present. There has been a contact with the antigen (pathogen resp. vaccine).

QUALITY ASSURANCE

Cross-Reactivity

Investigation of a sample panel with antibody activities to potentially cross-reacting parameters did not reveal evidence of false-positive results due to cross-reactions.

Sensitivity

The diagnostic sensitivity is defined as the probability of the assay of scoring positive in the presence of the specific analyte. It is 99.31% (95% confidence interval: 96.22% - 99.98%).

Interferences

Interferences with hemolytic, lipemic or icteric samples are not observed up to a concentration of 10 mg/mL hemoglobin, 5 mg/mL triglycerides and 0.5 mg/mL bilirubin.

Intra-assay and Inter-assay precision

The CV value of intra-assay precision was 2.62-5.63% and inter-assay precision was 4.25-8.97%.