

Monkey Leptin ELISA Kit

Enzyme Immunoassay for the quantification of Monkey Leptin in Monkey Serum and plasma

Catalog number: ARG82899

Package: 96 wells

For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	Page
INTRODUCTION	3
PRINCIPLE OF THE ASSAY	3
MATERIALS PROVIDED & STORAGE INFORMATION	4
MATERIALS REQUIRED BUT NOT PROVIDED	5
TECHNICAL HINTS AND PRECAUTIONS	5
SAMPLE COLLECTION & STORAGE INFORMATION	6
REAGENT PREPARATION	7
ASSAY PROCEDURE	9
CALCULATION OF RESULTS	11
EXAMPLE OF TYPICAL STANDARD CURVE	12
QUALITY ASSURANCE	12

MANUFACTURED BY:

Arigo Biolaboratories Corporation

Address: No. 22, Ln. 227, Gongyuan Rd., Hsinchu City 300, Taiwan

Phone: +886 (3) 562 1738

Fax: +886 (3) 561 3008

Email: info@arigobio.com

INTRODUCTION

This gene encodes a protein that is secreted by white adipocytes, and which plays a major role in the regulation of body weight. This protein, which acts through the leptin receptor, functions as part of a signaling pathway that can inhibit food intake and/or regulate energy expenditure to maintain constancy of the adipose mass. This protein also has several endocrine functions, and is involved in the regulation of immune and inflammatory responses, hematopoiesis, angiogenesis and wound healing. Mutations in this gene and/or its regulatory regions cause severe obesity, and morbid obesity with hypogonadism. This gene has also been linked to type 2 diabetes mellitus development. [provided by RefSeq, Jul 2008]

May function as part of a signaling pathway that acts to regulate the size of the body fat depot. An increase in the level of LEP may act directly or indirectly on the CNS to inhibit food intake and/or regulate energy expenditure as part of a homeostatic mechanism to maintain constancy of the adipose mass. [UniProt]

PRINCIPLE OF THE ASSAY

This assay employs the sandwich enzyme immunoassay technique. An antibody specific for Leptin has been pre-coated onto a microtiter plate. Standards or samples are pipetted into the wells and any Leptin present is bound by the immobilized antibody. After washing away any unbound substances, a biotin-conjugated antibody specific for Leptin is added to each well and incubate to bind to Leptin captured by the first antibody. Following a washing to remove unbound substances, streptavidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After washing away any unbound antibody-enzyme reagent, a substrate

solution (TMB) is added to the wells and color develops in proportion to the amount of Leptin bound in the initial step. The color development is stopped by the addition of acid and the intensity of the color is measured at a wavelength of 450nm ±2nm. The concentration of Leptin in the sample is then determined by comparing the O.D of samples to the standard curve.

MATERIALS PROVIDED & STORAGE INFORMATION

Store the unopened kit at 2-8°C. Use the kit before expiration date.

Component	Quantity	Storage information
Antibody-coated microplate	8 well X 12 strips	4°C
Standard	3 X 0.8 ng/Vial (Lyophilized)	4°C
Biotin-antibody conjugate concentrate	1 vials (Lyophilized)	4°C
HRP-Streptavidin conjugate concentrate	1 vial (55 μl)	4°C
20X PBS	25 ml	4°C
20X Assay Buffer	20 ml	4°C
TMB substrate	10.5 ml (Ready to use)	4°C (Protect from light)
STOP solution	5.5 ml (Ready to use)	4°C

MATERIALS REQUIRED BUT NOT PROVIDED

- Microplate reader capable of measuring absorbance at 450nm (optional: read at 540-570 nm as the reference wave length)
- Pipettes and pipette tips
- Deionized or distilled water
- Sterile 1 x PBS
- Automated microplate washer (optional)

TECHNICAL HINTS AND PRECAUTIONS

- Wear protective gloves, clothing, eye, and face protection especially while handling blood or body fluid samples.
- Store the kit at 4°C at all times.
- To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, a pilot experiment using standards and a small number of samples is recommended.
- Briefly spin down (6000Xg for 1 min) the Standards, Biotin-antibody conjugate and HRP-streptavidin conjugate before use.
- If crystals are observed in the 20X Assay Buffer, warm to RT until the crystals are completely dissolved.
- Ensure complete reconstitution and dilution of reagents prior to use.
- A standard curve should be generated for each set of samples assayed.
 Thorough mixing of standards at each of dilution steps is critical to acquire a normal standard curve.
- Brief vortex samples and diluted standards for 10 sec to mix well before add to the 96 well plate.

Monkey Leptin ELISA kit ARG82899

- All reagents should be mixed by gentle inversion or swirling prior to use.
 Do not induce foaming.
- Do not let strips dry, as this will inactivate active components in wells.
- Avoid using reagents from different batches.
- It is highly recommended that the standards, samples and controls be assayed in duplicates.
- HRP Conjugate contains enzyme, DO NOT mass up with Detection Antibody.
- The Stop Solution is an acid solution, handle with caution.
- Change pipette tips between the addition of different reagent or samples.

SAMPLE COLLECTION & STORAGE INFORMATION

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

<u>Cell Culture Supernatants</u> - Remove particulates by centrifugation for 10 min at 1500 x g at 4°C and aliquot & store samples at -20°C up to 1 month or-80°C up to 6 months. Avoid repeated freeze-thaw cycles.

<u>Serum</u>- Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Collect serum and assay immediately or aliquot & store samples at -20°C up to 1 month or -80°C up to 6 months. Avoid repeated freeze-thaw cycles.

<u>Plasma</u> - Collect plasma using EDTA or heparin as an anticoagulant. Centrifuge for 15 minutes at $1000 \times g$. within 30 minutes of collection. Collect the supernatants and assay immediately or aliquot and store samples at -20° C up to 1 month or -80° C up to 6 months. Avoid repeated freeze-thaw cycles.

Note:


- a) Do not use haemolytic, icteric or lipaemic specimens.
- b) Samples containing sodium azide should not be used in the assay.

REAGENT PREPARATION

- 1X PBS: Dilute 20X PBS into deionized distilled water to yield 1X PBS.
- **1X Assay Buffer:** Dilute **20X** Assay Buffer into **1X PBS** to yield 1X Assay buffer. The diluted 1X Assay Buffer can be stored at 4°C.
- 1x Biotin-antibody Conjugate: The lyophilized Biotin-antibody conjugate could be stored at 4°C to -20°C for up to 3 months. Centrifuge the vial for 1 min at 6000 x g to spin down the material prior to open the vial. Reconstitute the Biotin-antibody Conjugate with 200 μl of sterile 1 x PBS, vortex for 15 sec and keep the antibody in the vail for 5 min to completely dissolve. Centrifuge the vial for 1 min at 6000 x g before opening. Aliquot and store the antibody stock at -20°C until use. Avoid repeated freezethaw cycles.
 - If the entire 96-well plate is used, dilution of the 200 μ l of concentrated Biotin-Conjugate solution with 10.5 ml of 1X PBS to yield 1X Biotinantibody Conjugate working solution.
- **1X HRP-streptavidin conjugate:** Centrifuge the vial for 1 min at 6000 x g to spin down the material prior to open the vial. The stock vial includes 55 μl of HRP-streptavidin concentrate. Please confirm if the vial contains 55 μl of HRP-streptavidin concentrate before further dilution. If it is less than 55 μl, add sterile 1X PBS to reach 55 μl and vortex briefly for 10 sec. Make a 1:200 dilution of the concentrated HRP-streptavidin solution with 1X PBS (If the entire 96-well plate is used, add 53 μl of concentrated HRP-

- streptavidin solution into 10.5 ml 1X PBS and mix thoroughly prior to the assay). The rest of <u>undiluted</u> HRP-streptavidin Conjugate can be stored at 4°C for up to 3 months. **DO NOT FREEZE.**
- Sample: If the initial assay found samples contain Leptin higher than the highest standard, the samples can be diluted with 1 x Assay Buffer and then re-assay the samples. For the calculation of the concentrations this dilution factor has to be taken into account. The sample must be well mixed with 1 x Assay Buffer before assay.
 - (It is recommended to do pre-test to determine the suitable dilution factor).
- for up to 3 months. Centrifuge the vial for 1 min at 6000 x g to spin down the material prior to open the vial. Reconstitute the standard with **0.5 ml** of **1 x Assay Buffer** to yield a stock concentration of **1600 pg/ml**. Brief vortex the vials for 15 sec and keep the standard stock in the vail for 5 min to completely dissolve. Make sure the standard is dissolved completely and then centrifuge the vial for 1 min at 6000 x g before making serial dilutions. Aliquot and store the reconstituted standard at -20°C for up 2 days.

The 1 x Assay Buffer serves as zero standard (0 pg/ml), and the rest of the standard serial dilution can be diluted with 1X Assay Buffer as according to the suggested concentration below: 1600 pg/ml, 800 pg/ml, 400 pg/ml, 200 pg/ml, 100 pg/ml, 50 pg/ml, 25 pg/ml. Brief vortex the vials for 30 sec for each standard dilution steps to mix well.



Dilute Leptin standard as according to the table below:

Standard	Leptin Conc. (pg/ml)	μl of 1X Assay Buffer	μl of standard
S7	1600 pg/ml	0	500 (1600 pg/ml Stock)
S6	800 pg/ml	250	250 (S7)
S5	400 pg/ml	250	250 (S6)
S4	200 pg/ml	250	250 (S5)
S3	100 pg/ml	250	250 (S4)
S2	50 pg/ml	250	250 (S3)
S1	25 pg/ml	250	250 (S2)
S0	0	250	0

ASSAY PROCEDURE

All materials should be equilibrated to room temperature (RT) before use. Standards, samples and controls should be **assayed in duplicates.**

- Lift the plate cover from the top left and cover the wells that are not used.
 Brief vortex and then spin down the standards and samples for 10 sec to mix completely before applying to the plate.
- 2. Add $100 \,\mu l$ of standards, samples and zero controls (1X Assay Buffer) in duplicates into wells. Incubate for 2 hours at room temperature.
- 3. Aspirate each well and wash, repeating the process once for a **total two** washes. Wash by filling each well with 1× Assay Buffer (300 μl) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each is essential to good performance. After the last wash, remove any remaining Assay Buffer by aspirating, decanting or blotting against clean paper towels.
- 4. Add $100 \,\mu l$ of 1x Biotin-antibody Conjugate working solution to each well. Cover the plate and incubate 1 hour at room temperature.
- 5. Aspirate each well and wash as step 3.
- 6. Add $100\,\mu l$ of 1X HRP-Streptavidin solution to each well. Cover wells and incubate for 20 minutes at room temperature in dark.
- 7. Aspirate each well and wash as step 3 but wash for a total four washes.
- Add 100 μl of TMB Substrate Solution to each well. Incubate for 5-30 minutes (depending on signal) at room temperature in dark.
- 9. Add $50~\mu l$ of Stop~Solution to each well. Gently tap the plate to ensure thorough mixing.
- 10. Read the OD with a microplate reader at **450 nm** immediately. (Optional:


it is recommended to detect background signal by reading the signal at 540-570 nm as reference wavelength).

CALCULATION OF RESULTS

- 1. (Optional) Subtract the absorbance of the value reading at 540-570 nm from the value reading at 450 nm.
- 2. Calculate the average absorbance values for each set of standards, controls and patient samples.
- 3. Using log-log, semi-log or linear graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
- 4. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
- 5. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
- 6. arigo provides GainData®, an in-house development ELISA data calculator, for ELISA data result analysis. Please refer our GainData® website for details. (https://www.arigobio.com/elisa-analysis)
- 7. If the samples have been diluted, the concentration read from the standard curve must be further converted by the appropriate dilution factor according to the sample preparation procedure as described above.

EXAMPLE OF TYPICAL STANDARD CURVE

The following data is for demonstration only and cannot be used in place of data generations at the time of assay.



QUALITY ASSURANCE

Sensitivity

The minimum detectable dose (MDD) of Monkey Leptin ranged from 25-1600 pg/ml. The mean MDD was 13 pg/ml.

Specificity

This assay recognizes natural and recombinant Monkey Leptin. No significant cross-reactivity or interference with the factors below was observed:

Not react with following recombinant Monkey proteins: Adiponectin, ApoAl, BMP1, BMP2, BMP3, BMP4, BMP5, BMP7, CCL2, CCL4, CCL5, CRP, HSP27, HGF, IL-1 beta, IL-1RA, IL-2, IL-4, IL-5, IL-6, sIL-6R, IL-8, IL-10, IL-12, IL-15, IL-17C, IL-21, IL-23, IFN γ , MMP-2, MMP-9, IL2R, PDGF, serpin E1, TGF β 1, TGF β 2, TGF β 3, TLR1, TLR2, TLR3, TLR9, TNF- α , TNF RI, TNF RII, VEGF, VEGF R1.

Intra-assay and Inter-assay precision

The CV values of intra-assay was 7% and inter-assay was 10%.