



Cas9 ELISA Kit

ARG83379 Cas9 ELISA Kit is an Enzyme Immunoassay kit for the quantification of Cas9 in cell lysate and tissue lysate

Catalog number: ARG83379

Package: 96 wells

For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	Page
INTRODUCTION	3
PRINCIPLE OF THE ASSAY	3
MATERIALS PROVIDED & STORAGE INFORMATION.....	4
MATERIALS REQUIRED BUT NOT PROVIDED.....	4
TECHNICAL NOTES AND PRECAUTIONS	5
SAMPLE COLLECTION & STORAGE INFORMATION	6
REAGENT PREPARATION	7
ASSAY PROCEDURE	8
CALCULATION OF RESULTS.....	9
EXAMPLE OF TYPICAL STANDARD CURVE.....	10

MANUFACTURED BY:

Arigo Biolaboratories Corporation

Address: 9F.-7, No. 12, Taiyuan 2nd St., Zhubei City,

Hsinchu County 302082, Taiwan

Tel: +886-3-6221320

Fax: +886-3-6221321

Email: info@arigobio.com

INTRODUCTION

Cas9 is a 160 kilodalton protein which plays a vital role in the immunological defense of certain bacteria against DNA viruses and plasmids, and is heavily utilized in genetic engineering applications. Its main function is to cut DNA and thereby alter a cell's genome.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. An antibody specific for CAS9 has been pre-coated onto a microtiter plate. Standards or samples are pipetted into the wells and any CAS9 present is bound by the immobilized antibody. After washing away any unbound substances, a biotin-conjugated antibody specific for CAS9 is added to each well and incubate. Following a washing to remove unbound substances, streptavidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After washing away any unbound antibody-enzyme reagent, a substrate solution (TMB) is added to the wells and color develops in proportion to the amount of CAS9 bound in the initial step. The color development is stopped by the addition of acid and the intensity of the color is measured at a wavelength of $450\text{nm} \pm 2\text{nm}$. The concentration of CAS9 in the sample is then determined by comparing the O.D of samples to the standard curve.

Cas9 ELISA Kit ARG83379

MATERIALS PROVIDED & STORAGE INFORMATION

Upon received, store Standard at -80°C, 500X Cas9 Antibody at -20°C, Store other component at 2-8°C at all times. Use the kit before expiration date.

Component	Quantity	Storage information
Antibody Coated Microplate	8 x 12 strips	4°C.
Standard	1 vial (50 µL; 5 µg/mL)	-80°C.
1000X Cas9 Antibody	1 vial (10 µL)	-20°C
1000X HRP-Streptavidin Solution	1 vial (20 µL)	4°C
Diluent Buffer	50 mL	4°C
10X Wash Buffer	100 mL	4°C
TMB Substrate	12 mL	4°C (Protect from light)
Stop Solution	12 mL	4°C

MATERIALS REQUIRED BUT NOT PROVIDED

- Microplate reader capable of reading at O.D. 450 nm
- Centrifuge and centrifuge tube
- Deionized or Distilled water
- Pipettes, pipette tips and Multichannel micropipette reservoir
- Automated microplate washer (optional)

TECHNICAL NOTES AND PRECAUTIONS

- Wear protective gloves, clothing, eye, and face protection especially while handling blood or body fluid samples.
- Upon received, store Standard at -80°C, 500X Cas9 Antibody at -20°C, Store other component at 2-8°C at all times. Use the kit before expiration date.
- Prior to beginning the assay procedure, bring all reagents and required number of strips to room temperature.
- Briefly spin down the all vials before use.
- If crystals are observed in the 10X Wash Buffer, warm to 37°C until the crystals are completely dissolved.
- Minimize lag time between wash steps to ensure the plate does not become completely dry during the assay.
- Ensure complete reconstitution and dilution of reagents prior to use.
- Take care not to contaminate the TMB Substrate. Do not expose the TMB solution to glass, foil or metal. Do NOT return leftover TMB Substrate to bottle. Do NOT contaminate the unused TMB Substrate. If the solution is blue before use, DO NOT USE IT.
- Change pipette tips between the addition of different reagent or samples.
- Taping the well strips together with lab tape can be done as an extra precaution to avoid plate strips from coming loose during the procedure.
- Include a standard curve each time the assay is performed.
- Run both standards and samples in at least duplicates.

SAMPLE COLLECTION & STORAGE INFORMATION

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

Cells or tissues: Sonicate or homogenize sample in Lysis Buffer such as RIPA buffer (25 mM Tris-HCl pH 7.6, 150 mM NaCl, 1% NP-40, 1% sodium deoxycholate, 0.1% SDS) and centrifuge at 10,000 x g for 10 minutes at 4°C. Assay immediately or store samples at -80°C for up to three months. Dilute samples in PBS containing 0.1% BSA as needed.

REAGENT PREPARATION

- **1X Wash Buffer:** Dilute **10X** Wash Buffer into distilled water to yield 1X Wash Buffer.
- **1X Cas9 Antibody:** 10 minutes before use, dilute **1000X** Cas9 Antibody into Diluent Buffer to yield **1X** Cas9 Antibody.
- **1X HRP-Streptavidin Solution:** 10 minutes before use, dilute **1000X** HRP-Streptavidin Solution into Diluent Buffer to yield **1X** HRP-Streptavidin Solution. Keep diluted HRP-Streptavidin Solution in dark before use.
- **Standard:** Centrifuge the un-reconstituted standard at 6000 x g for 1 minute to bring down the material prior to open the vial. The Diluent Buffer serves as zero standard (0 pg/ml), and the rest of the standard serial dilution can be diluted with Diluent Buffer. Diluted the standard as below:

Standard tube	Cas9 (ng/mL)	Diluent Buffer (μL)	Standard (μL)
S1	100	490	10 (50 μg/mL Stock)
S2	50	250	250 of S1
S3	25	250	250 of S2
S4	12.5	250	250 of S3
S5	6.25	250	250 of S4
S6	3.125	250	250 of S5
S7	1.562	250	250 of S6
S0	0	250	0

Note: Working standard should be prepared immediately prior to use.

ASSAY PROCEDURE

All materials should be equilibrated to room temperature (RT, 20-25°C) before use. Standards and samples should be assayed in duplicates.

1. Add **100 µL** of **Samples, Standard** into respective wells of the 96-well plate.
2. Cover the plate and incubate for **1 hour** at **RT**.
3. Aspirate each well and wash, repeating the process 2 time for a **total 3 washes**. Wash by filling each well with **1X Wash Buffer (250 µL)** using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating, decanting or blotting against clean paper towels.
4. Add **100 µL** of **1X Cas9 Antibody** to each well.
5. Cover the plate and incubate for **1 hour** at **RT**.
6. Aspirate each well and **wash plate as step 3**.
7. Add **100 µL** of **1X HRP-Streptavidin Solution** to each well.
8. Cover the plate and incubate for **1 hour** at **room temperature** in the dark.
9. Aspirate each well and **wash plate as step 3**.
10. Add **100 µL** of **TMB Substrate** in each well.
11. Incubate for **5-30 mins** at **room temperature** in the dark.
12. Add **100 µL** of **Stop Solution** to each well to stop the reaction.
13. Read the absorbance with a plate reader at **O.D. 450 nm**. It is recommended reading the absorbance within 10 minutes after adding the stop solution.


CALCULATION OF RESULTS

1. Calculate the average absorbance values for each set of standards, control and samples.
2. Using log-log, semi-log or linear graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. arigo provides GainData®, an in-house development ELISA data calculator, for ELISA data result analysis. Please refer our GainData® website for details. (<https://www.arigobio.com/elisa-analysis>)
6. If the samples have been diluted, the concentration read from the standard curve must be further converted by the appropriate dilution factor according to the sample preparation procedure as described above.

Cas9 ELISA Kit ARG83379

EXAMPLE OF TYPICAL STANDARD CURVE

The following figures demonstrate typical results with the Cas9 ELISA Kit. One should use the data below for demonstration only and cannot be used in place of data generations at the time of assay.



QUALITY ASSURANCE

Sensitivity

1.0 ng/mL

Assay Range

1.56- 100 ng/mL