

Glutelin Assay Kit

ARG83403 Glutelin Assay Kit can be used to measure Glutelin in tissue extracts and powder.

Catalog number: ARG83403

Package: 96 wells

For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	Page
INTRODUCTION	3
PRINCIPLE OF THE ASSAY	3
MATERIALS PROVIDED & STORAGE INFORMATION	3
MATERIALS REQUIRED BUT NOT PROVIDED	4
TECHNICAL HINTS AND PRECAUTIONS	4
SAMPLE COLLECTION & STORAGE INFORMATION	4
REAGENT PREPARATION	5
ASSAY PROCEDURE	6
CALCULATION OF RESULTS	6
EXAMPLE OF TYPICAL RESULT	7

MANUFACTURED BY:

Arigo Biolaboratories Corporation

Address: No. 22, Ln. 227, Gongyuan Rd., Hsinchu City 300, Taiwan

Phone: +886 (3) 562 1738

Fax: +886 (3) 561 3008

Email: info@arigobio.com

INTRODUCTION

Glutelins are a class of prolamin proteins found in the endosperm of certain seeds of the grass family. They constitute a major component of the protein composite collectively referred to as gluten. Glutenin is the most common glutelin, as it is found in wheat and is responsible for some of the refined baking properties in bread wheat. The glutelins of barley and rye have also been identified. Glutelins are the primary protein form of energy storage in the endosperm of rice grains.

PRINCIPLE OF THE ASSAY

The Glutelin Assay Kit can measure Glutelin in plant samples. The increase in absorbance at 595 nm is directly proportional to the Hordein content.

MATERIALS PROVIDED & STORAGE INFORMATION

Component	Quantity	Storage	
Microplate	1 X 96-well plate		
Standard	1 vial (lyophilized)	-20°C	
Assay Buffer I	2 X 30 ml (ready to use)	4°C	
Assay Buffer II	2 X 30 ml (ready to use)	4°C	
Assay Buffer III	2 X 30 ml (ready to use)	4°C	
Assay Buffer IV	2 X 30 ml (ready to use)	4°C	
Reagent Dye	20 ml	4°C	

MATERIALS REQUIRED BUT NOT PROVIDED

- Microplate reader capable of measuring absorbance at 595 nm
- Pipettes and pipette tips
- Deionized or distilled water

TECHNICAL HINTS AND PRECAUTIONS

- Wear protective gloves, clothing, eye, and face protection especially while handling blood or body fluid samples.
- Standard store at -20°C, all other component store at 4°C.
- Briefly spin down the reagents before use.
- It is highly recommended that the standards and samples be assayed in at least duplicates.
- Change pipette tips between the addition of different reagent or samples.

SAMPLE COLLECTION & STORAGE INFORMATION

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

<u>Tissue -</u> Weigh out 0.05 g tissue, homogenize with 0.5 ml Assay Buffer I on ice, transfer it to centrifuge tube and mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the supernatant; then add 0.5 ml Assay Buffer II into the tube, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the supernatant; then add 0.5 ml Assay Buffer III into the tube, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the

supernatant; then add 0.5 ml Assay Buffer IV into the tube, mix on a lab rotator

Glutelin Assay Kit ARG83403

for 30 minutes, centrifuged at 10000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

<u>Powder -</u> Weigh out 0.05 g powder, add 0.5 ml Assay Buffer I to dissolve, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the supernatant; then add 0.5 ml Assay Buffer II into the tube, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the supernatant; then add 0.5 ml Assay Buffer III into the tube, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, discard the supernatant; then add 0.5 ml Assay Buffer IV into the tube, mix on a lab rotator for 30 minutes, centrifuged at 10000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

REAGENT PREPARATION

- Standard: Add 1 ml of distilled water to yield 2 mg/ml standard. Perform 2fold serial dilution of the top standards to make the standard curve.
- Sample: If the measuring absorbance of samples is higher than the standard, dilute the samples with Distilled water before assay and assay again. For the calculation of the activity this dilution factor has to be taken into account.

ASSAY PROCEDURE

Standards and samples should be assayed in at least duplicates.

- 1. Sample wells: Add **10 μl Sample** into Sample wells.
- 2. Standard wells: Add 10 µl Standard into Standard wells.
- 3. Add 200 µl Reagent Dye to each wells.
- 4. Mix well. Incubate at **RT** for **2 min**. Read the OD at **595nm**

Reagent	Sample	Standard	Blank	
Sample	10 μΙ	-	-	
Standard	-	10 μΙ	-	
Distilled water	-		10 μΙ	
Reagent Dye	200 μΙ	200 μΙ	200 μΙ	
Mix well. Incubate at RT for 2 min. Read the OD at 595nm				

CALCULATION OF RESULTS

- 1. Calculate the average absorbance values for each set of samples, standard and blank.
- 2. Calculation:

A. Definition:

C_{Standard}: the standard concentration, 2 mg/ml;

W: the weight of sample, g;

 V_{Sample} : the volume of reaction sample, 10 μ l = 0.01 ml;

 V_{standard} : the volume of standard sample, 10 µl = 0.01 ml;

 V_{assay} : the volume of Assay Buffer III, 500 μ l = 0.5 ml.

Glutelin Assay Kit ARG83403

B. Formula:

a). According to the weight concentration of sample

3. Detection range:

The detection range is from 0.02 mg/ml - 2 mg/ml.

4. If the samples have been diluted, the calculated activity must be further converted by the appropriate dilution factor according to the sample preparation procedure as described above.

EXAMPLE OF TYPICAL RESULT

The following data is for demonstration only and cannot be used in place of data generations at the time of assay. Please note this data is for demonstration only and serial diluted standards are not necessary for this kit.

